

**King, Valerie A.**

---

**From:** Kimberly Prchal [Kimberly.Prchal@mdsps.com]  
**Sent:** Wednesday, May 26, 2004 4:09 PM  
**To:** Mohamadi.Sarkar@pmusa.com; Steve Hambek  
**Cc:** valerie.a.king@pmusa.com; Nancy Wang; Yuli Xie; Henry Cheng; Susan Mitmesser; Jacqueline Mowitz; Angela Stiens; shixia.feng@pmusa.com; robin.d.kinser@pmusa.com; barbara.k.zedler@pmusa.com  
**Subject:** Re: SCoR2003/11/01/03 Creatinine Re-Assay Results  
**Sensitivity:** Confidential

Steve,

Please create NTF describing MDS process for entering data into the spreadsheet and the reasoning for the calculation error.

Thanks  
Kim

Mohamadi,

Attached are the results from the re-assay of creatinine results for the SCoR2003/01/03 (AA02243) study. As Steve mentions in the attached e-mail, there seems to have been a calculation error in the original results.

The recalculated values are the original values with the correct calculations which brings them very close to the reassayed values. There seem to be no assay problems with the original values, just the wrong factor in calculation. If need be, Steve would be able to create a Note to File describing the error. Please advise.

Please review and let MDS know which data you would like to use.

Thanks  
Kim

Kimberly Prchal  
Senior Project Manager  
MDS Pharma Services  
621 Rose Street  
Lincoln, NE 68502 USA  
kimberly.prchal@mdsps.com  
P:402-437-4773  
C:402-610-0190